



CERTIFICATE OF ANALYSIS

pH buffer solution $4.00 \pm 0.05 @ 25^{\circ}\text{C}$

Order No: 901.12.00500.02

Batch No: 901.12.03-E10.21


Date of analysis: 10/05/2021

Minimum shelf life: 10/05/2023

Test	Mean measured value
pH value	3.992

The pH value of this standard solution was determined electrochemically using combined glass electrode under standard laboratory condition. The uncertainty of measurement has been calculated not to exceed ± 0.02 pH (coverage factor =2 for 95% confidence level).

The certified value has been compared with a reference pH standard solution which is directly traceable to primary standard NIST SRM 185i (National Institute of Standards and Technology, USA).

QC. Manager:	Qualified Person:
Status: 	Date: 10.05.2021

Document code : QC.PC.WIN.005/COA.03

